


9500 Fluid Warmer Temperature Chart

Flowrate, ml/hr	Temperature, ° F
100	104
110	104
120	104
130	104
140	104
150	104
183	103.1
254	101.3
300	99.5
338	98.6



Note: The temperatures shown on the chart are exit temperatures from the fluid warmer.
Fluid temperature at the injection site depends on distance, ambient temperature, and flowrate.